

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0438 Accredited to ISO/IEC 17025:1999	European Instruments	
	Issue No: 007 Issue date: 30 June 2006	
	Shotover Kilns Old Road Headington Oxford OX3 8ST	Contact: Mr A P Janes Tel: +44 (0)1865-750375 Fax: +44 (0)1865-769985 E-Mail: balances@euroinst.co.uk Website: www.euroinst.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Shotover Kilns Old Road Headington Oxford OX3 8ST	Local contact Contact: Mr A P Janes Tel: +44 (0)1865-750375 Fax: +44 (0)1865-769985	Mass Mass (weights) Volume Volumetric equipment (pipettes) A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Any premises	Mass Weighing machines (non-automatic) Volume Volumetric equipment (pipettes)	Site



0438
Accredited to
ISO/IEC 17025:1999

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

European Instruments

Issue No: 007 Issue date: 30 June 2006

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
MASS (See Notes 1 & 2)	Nominal value (g) 20 000 10 000 5 000 2 000 1 000 500 200 100 50 20 10 5 2 1 0.5 0.2 0.1 0.05 0.02 0.01 0.005 0.002 0.001	(mg) 60 10 5 2 1 0.5 0.1 0.05 0.03 0.025 0.020 0.015 0.012 0.010 0.008 0.006 0.005 0.004 0.003 0.002 0.002 0.002	NOTES 1. Calibrations can be given in other units as required. 2. Intermediate values of weights can be calibrated with an uncertainty not less than that interpolated from the next higher and next lower nominal values.	A
NON-AUTOMATIC WEIGHING MACHINES	Up to 1350 kg	Uncertainties quoted will depend on the performance of the weighing machine under calibration, and will not be less than the uncertainty of calibration of the weights used for the calibration.	1. Weights are available in OIML Class E2 from 1 mg to 10 kg, Class F1 from 1 g to 10 kg, Class F2, 2 x 10 kg and Class M1 at 5, 10 and 20 kg to a total of 1350 kg.	Site
VOLUME of liquids (See Note 3)	From 1 µL to 20 µL Above 20 µL to 100 µL Above 100 µL to 200 µL Above 200 µL to 500 µL Above 0.5 mL to 1 mL Above 1 mL to 2 mL Above 2 mL to 5 mL Above 5 mL to 10 mL	0.09 µL 0.3 µL 0.8 µL 2.6 µL 0.005 mL 0.013 mL 0.022 mL 0.075 mL	3. For water delivered from piston and/or plunger operated volumetric apparatus using procedures agreed with UKAS.	A, Site
END				