

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

	<b>European Instruments</b>	
	<b>Issue No:</b> 007 <b>Issue date:</b> 03 September 2004	
	<b>Shotover Kilns</b> Old Road Headington Oxford OX3 8ST	<b>Contact:</b> Mr A P Janes <b>Tel:</b> +44 (0)1865-750375 <b>Fax:</b> +44 (0)1865-769985 <b>E-Mail:</b> <a href="mailto:balances@euroinst.co.uk">balances@euroinst.co.uk</a> <b>Website:</b> <a href="http://www.euroinst.co.uk">www.euroinst.co.uk</a>

### SUMMARY OF ACCREDITATION

Calibration performed on site

**MASS**

Weighing machines (non-automatic)

**VOLUME**

Volumetric equipment



Site calibration

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

### European Instruments

**Issue No:** 007    **Issue date:** 03 September 2004

#### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
NON-AUTOMATIC WEIGHING MACHINES	Up to 1350 kg	Uncertainties quoted will depend on the performance of the weighing machine under calibration, and will not be less than the uncertainty of calibration of the weights used for the calibration.	NOTES 1 Weights are available in OIML Class E2 from 1 mg to 10 kg, Class F1 from 1 g to 10 kg, Class F2, 2 x 10 kg and Class M1 at 5, 10 and 20 kg to a total of 1350 kg.
VOLUME of liquids (See Note 2)	From 1 $\mu$ L to 20 $\mu$ L Above 20 $\mu$ L to 100 $\mu$ L Above 100 $\mu$ L to 200 $\mu$ L Above 200 $\mu$ L to 500 $\mu$ L Above 0.5 mL to 1 mL Above 1 mL to 2 mL Above 2 mL to 5 mL Above 5 mL to 10 mL	0.09 $\mu$ L 0.3 $\mu$ L 0.8 $\mu$ L 2.6 $\mu$ L 0.005 mL 0.013 mL 0.022 mL 0.075 mL	2 For water delivered from piston and/or plunger operated volumetric apparatus using procedures agreed with UKAS.
END			